

## Material Safety Data Sheet – InSight Urinalysis Analyser

### Section 1 – Product and Company Identification

**Manufacturer:** Woodley Equipment Company Ltd.

Old Station Park Buildings, St. John Street, Horwich, Bolton, BL6 7NY, UK

Tel: +44 (0) 1204 669033

Email: sales@woodleyequipment.com

**Product Name:** InSight Urinalysis Veterinary Urine Analyser.

**Chemical Family or Product Type:** Non-medical devices.

### Section 2 – Hazard Identification

**Hazard Designation:** Void.

**Information Pertaining to Particular Dangers for Man and Environment:** Void.

**Classification System:** Non-medical devices.

### Section 3 – Special Safety Measures

Not applicable.

### Section 4 – Storage Environment

**Environment Temperature:** -20°C to +55°C.

**Comparative Humidity:** ≤95% RH.

The packed instrument should be stored in a ventilated room. Do not store the instrument with toxicant, injurant and corrosive.

### Section 5 – Fire Fighting Measures

**Extinguishing Media:** Use any means which is suitable for the surrounding fire.

**Protective Equipment:** No special measures required.

## Section 6 – Transport Information

**Land Transport ADR/RID and GGVS/GGVE (Cross-border/Domestic)**

**ADR/RID-GGVS/E Class:** Not regulated.

**Conditions to Avoid:** Excessive heat or cold, direct light, high humidity.

Transport the instrument with the utmost care and avoid moisture, sunlight and collision.

**Transportation Conditions:** -20°C to +55°C, ≤95% RH.

## Section 7 – Exposure Controls

### Urine Samples

**Handling:** Keep samples away from direct sunlight.

**Protective Clothing:** Wear protective gloves to avoid exposure to urine samples.

### Test Strips

Exposing the test strips to air for a long time will deteriorate the composition of the test strips.

Do not touch the reagent pads on the test strip.

## Section 8 – Operating Environment

**Operating Temperature:** +5°C to +40°C (Optimum Operating Temperature +23°C to +28°C).

**Measuring Temperature:** +15°C to +30°C.

**Comparative Humidity:** ≤80% RH.

**Barometric:** 75kpa – 106kpa.

## Section 9 – Safety Power

**Voltage:** AC 100 – 240 (1±10%) V

**Frequency:** 50/60 Hz

## Section 10 – Disposal Considerations

Put on protective gloves to avoid exposure to pathogenic microbes.

Dispose of used samples, test strips and gloves according to local regulations.

## Section 11 – Handling

The instrument must be handled with care. Rough handling may disturb internal calibration and/or cause other malfunctions. An error code will appear on the display if there is something wrong with the instrument and the error code will also be printed out if the printer is attached. The list of error codes, possible causes and recommended solutions can be found in the InSight Urinalysis Operator's Manual. If the error code persists, contact Woodley Equipment or your local distributor.

## Section 12 – Other Information

To the best of our knowledge, the information provided herein is accurate but does not purport to be all inclusive. It is intended to provide a general guidance in terms of safe handling, storage and disposal of materials. Woodley Equipment Company thus assumes no liabilities for any damage or loss resulting from handling or from contact with this product. Contact Woodley Equipment Company if additional information is needed.

**Prepared By:** Woodley Equipment Company Ltd.

Old Station Park Buildings, St. John Street, Horwich, Bolton, BL6 7NY, UK

Tel: +44 (0) 1204 669033

Email: [sales@woodleyequipment.com](mailto:sales@woodleyequipment.com)

**Date:** 9<sup>th</sup> July 2024